



90 Sargent Drive • New Haven, Connecticut 06511-5966
203-401-6743 or 877-894-5773 (Telephone) • Website: www.rwater.com

Laboratory Certification IDs

CT: PH-0411, NH: 2239, NY: 11867, PA: 68-05519, RI: LAO00339, MA: M-CT004

See website, www.rwalab.com/rwa-lab-certifications, for certified analyte list.

Client: Bill Cookerly and Noreen Jordan
Randolph-Holbrook Joint Water
50 N. Franklin Street

Holbrook, MA 02343
617-470-1624 cookertyb@gmail.com
781-767-1800 NJordan
NJordan@Holbrookmassachusetts.us

ANALYTICAL REPORT

Project: FEE-RANDOLPHHOLBROOK-22-000011

Report Date: 12/13/2022

This Laboratory is in compliance with the NELAP requirements of procedures used except where indicated.
This report contains results for the analysis tested, under the sampling conditions described on the Chain Of Custody (COC), as received by the laboratory. This report is incomplete unless all pages indicated in the pagination at the bottom of the page are included.

The COC form has been scanned to accompany the analytical report and is an exact copy of the original.

If you are the client above and have any questions concerning this testing, please do not hesitate to contact RWA Client Services at (203) 401-6743 or (877) 894-5773. The contents of this report cannot be discussed with anyone other than the client listed above without their written consent.

A handwritten signature in black ink, reading "Gary M. Alexander".

Gary M. Alexander, Team Lead
Technical Representative

Sample ID: 300255004

Customer ID: Finished Water

Collection Date: 10/05/2022 12:00

PWS ID# / LOC ID#: 4244001 / 10296

Project: FEE-RANDOLPHHOLBROOK-22-000011

Analyte	CAS#	Results	MDL	RL	Units	Dilution	Qualifier	Method	Date Time/ Analyzed	Analyst
PFBS	375-73-5	2.16	0.89	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFHxA	307-24-4	4.19	0.68	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFHpA	375-85-9	2.34	0.62	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFHxS	355-46-4	3.60	0.90	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFOA	335-67-1	5.18	0.72	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFOS	1763-23-1	8.13	0.89	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFNA	375-95-1	1.13	0.63	2.0	ng/L	1.0	J	EPA537	10/18/22 1915	CSS
PFDA	335-76-2	ND	0.72	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFUnA	2058-94-8	ND	0.91	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
NMeFOSAA	2355-31-9	ND	1.19	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
NEtFOSAA	2991-50-6	ND	1.03	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFDaA	307-55-1	ND	0.83	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFTTrDA	72629-94-8	ND	0.90	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFTA	376-06-7	ND	0.97	2.0	ng/L	1.0		EPA537	10/18/22 1915	CSS
PFAS6 (MassDEP)		19.25	2.00	2.00	ng/L	1.0				
Surrogates		Results		Recovery Limits		Pass/Fail				
13C-PFHxA (SUR) % Recovery		106.22		70 - 130		Pass				
13C-PFDA (SUR) % Recovery		92.08		70 - 130		Pass				
d5-NEtFOSAA (SUR) % Recovery		83.67		70 - 130		Pass				

Sample Extraction Data:

Lab Number (Field ID)	Prep Method	Batch	Final (mL)	Date
300255004	PFAS_537	537_EXT-221014-1	250	10/14/2022

Sample ID: 300255002

Customer ID: Raw Water

Collection Date: 10/05/2022 12:00

PWS ID# / LOC ID#:

Project: FEE-RANDOLPHHOLBROOK-22-000011

Analyte	CAS#	Results	MDL	RL	Units	Dilution	Qualifier	Method	Date Time/ Analyzed	Analyst
PFBS	375-73-5	1.98	0.89	2.0	ng/L	1.0	J	EPA537	10/18/22 1850	CSS
PFHxA	307-24-4	3.75	0.68	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFHpA	375-85-9	2.21	0.62	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFHxS	355-46-4	2.23	0.90	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFOA	335-67-1	4.85	0.72	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFOS	1763-23-1	7.44	0.89	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFNA	375-95-1	1.04	0.63	2.0	ng/L	1.0	J	EPA537	10/18/22 1850	CSS
PFDA	335-76-2	ND	0.72	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFUnA	2058-94-8	ND	0.91	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
NMeFOSAA	2355-31-9	ND	1.19	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
NEtFOSAA	2991-50-6	ND	1.03	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFDoA	307-55-1	ND	0.83	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFTTrDA	72629-94-8	ND	0.90	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFTA	376-06-7	ND	0.97	2.0	ng/L	1.0		EPA537	10/18/22 1850	CSS
PFAS6 (MassDEP)		16.73	2.00	2.00	ng/L	1.0				
Surrogates		Results		Recovery Limits		Pass/Fail				
13C-PFHxA (SUR) % Recovery		96.23		70 - 130		Pass				
13C-PFDA (SUR) % Recovery		90.19		70 - 130		Pass				
d5-NEtFOSAA (SUR) % Recovery		81.85		70 - 130		Pass				

Sample Extraction Data:

Lab Number (Field ID)	Prep Method	Batch	Final (mL)	Date
300255002	PFAS_537	537_EXT-221014-1	250	10/14/2022